

## Certificate of Conformity

Part No.	8050H-500 8050H-500S 8050H-1000 8050H-1000S
Description	Liqui Thief
Material	HDPE (High Density Polyethylene) Silicone Rubber 'O' Ring
Date	20-Oct-08

Dear Sir/Madam,

We confirm that the samplers (as detailed above) are manufactured exclusively from:

- HDPE that conforms to FDA CFR 177.1520
- Silicone rubber (virgin) that conforms to FDA CFR 177.2600

The samplers are assembled and packed in a class 100,000 medical cleanroom.

Ultra-sonic welding is used adhere the components.

For and on behalf of  
Sampling Systems,



Jonathan Fenner