

CERTIFICATE OF IRRADIATION

Number: **B** 0379134 00

SAM002

Page 1

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: S08-10014
Date Rec'd 4/08/08
Date: 13/08/08

| ITEM CODE ISOTRON BATCH | ITEM SPECIFICATION | QTY | ADDITIONAL DETAILS |
|----------------------------|-------------------------------------|-----|--|
| B2SAM0020006 B208080242 | 8085A-100S SACK SAMPLERS 25-45kGy | 3 | B/No. 74631 ACTUAL..... <u>35.4</u>kGy DOSE |
| B2SAM0020008 B208080242 | 8055H-500S DISP VISCOTHIEF 25-45kGy | 10 | B/No. 190406 ACTUAL..... <u>35.9</u>kGy DOSE |
| B2SAM0020013 B208080242 | 8060H-1005S DISP POWDER 25-45kGy | 9 | B/No. 187641 ACTUAL..... <u>35.8</u>kGy DOSE |
| B2SAM0020014 B208080242 | 8060H-1010S DISP POWDER 25-45kGy | 10 | B/No. 185870 ACTUAL..... <u>35.8</u>kGy DOSE |
| B2SAM0020015 B208080242 | 8050H-500S DISP LIQUITHIEF 25-45kGy | 10 | B/No. 194353 ACTUAL..... <u>35.9</u>kGy DOSE |
| B2SAM0020019 B208080242 | 8060H-1020S DISP POWDER 25-45kGy | 17 | B/No. 193185 |
| | | | Continued.. 1 |

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED

SAM002

Page 2

Sampling Systems Ltd
 2 Forge Mills Park
 Station Road
 Coleshill
 Warwickshire



Isotron Limited
 Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

B46 1HT
 Cust. Ref: S08-10014
 Date Rec'd 4/08/08
 Date: 13/08/08

| ITEM CODE ISOTRON BATCH | ITEM SPECIFICATION | QTY | ADDITIONAL DETAILS |
|----------------------------|--------------------------------------|-----|---|
| | | | ACTUAL..... 35.8kGy DOSE |
| B2SAM0020027 B208080242 | 8045A-1000S STERILE SCOOPS 25-45kGy | 180 | B/No.195896 ACTUAL..... 36.5 - 36.9kGy DOSE |
| B2SAM0020028 B208080242 | 8045A-60S DISP PHARMA SCOOP 25-45kGy | 20 | B/No.194352 ACTUAL..... 36.5kGy DOSE |
| B2SAM0020029 B208080242 | 8051H-900S JUMBO PIPETTE 25-45kGy | 27 | B/No.195975 ACTUAL..... 33.7 - 34.1kGy DOSE |
| B2SAM0020040 B208080242 | 5055G-50TS GLASS PIPETTES 25-45kGy | 8 | N/A ACTUAL..... 34.6kGy DOSE |
| B2SAM0020042 B208080242 | 8095A-150S POWDER SPATULA 25-45kGy | 40 | B/No.196002 ACTUAL..... 36.0 - 37.5kGy DOSE |
| | | | Continued.. 2 |

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drake*

For and on behalf of ISOTRON LIMITED

SAM002

Page 3

Sampling Systems Ltd
 2 Forge Mills Park
 Station Road
 Coleshill
 Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

B46 1HT

Cust. Ref: S08-10014
 Date Rec'd 4/08/08
 Date: 13/08/08

| ITEM CODE ISOTRON BATCH | ITEM SPECIFICATION | QTY | ADDITIONAL DETAILS |
|----------------------------|--------------------|-----|-------------------------------------|
| B2SAM002VAL | VALIDATION | | SINGLE VALIDATION ON POWDER SPATULA |
| B2SAM002MIS | MISCELLANEOUS | | 3 PALLETS |
| B2SAM002MIS | MISCELLANEOUS | | LABOUR COST FOR SPLITTING DOWN PLTS |
| B2SAM002MIS | MISCELLANEOUS | | HANDLING CHARGE |
| | Total | 334 | Last Page 3 of 3 |

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of ISOTRON LIMITED