

SAM002

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: S08-10079
Date Rec'd 10/09/08
Date: 15/09/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020006 B208090297	8085A-100S SACK SAMPLERS 25-45kGy	15	15 X B/No.77060 ACTUAL 32.5kGy DOSE
B2SAM0020008 B208090297	8055H-500S DISP VISCOTHIEF 25-45kGy	10	10 X B/No.195245 ACTUAL 34.3kGy DOSE
B2SAM0020016 B208090297	8050H-1000S DISP LIQUITHIEF 25-45kGy	40	10 X B/No.195245 30 X B/No.196081 ACTUAL 34.4 - 34.9kGy DOSE
B2SAM0020021 B208090297	8095A-150S POWDER SPATULA 25-45kGy	40	40 X B/No.196002 ACTUAL 34.7 - 35.6kGy DOSE
B2SAM0020023 B208090297	8063H-500S DISP POWDER 25-45kGy	52	52 X B/No.196518 ACTUAL 34.4 - 35.9kGy DOSE
B2SAM0020043	8045A-2500S DISP SCOOP 25-45kGy	59	54 X B/No.196698

Continued.. 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *E. Drice*

For and on behalf of ISOTRON LIMITED

CERTIFICATE OF IRRADIATION

Number: **B** 0382721 00

SAM002

Page 2

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: S08-10079
Date Rec'd 10/09/08
Date: 15/09/08

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B208090297			5 X B/No.196707 ACTUAL.....35.5 - 36.6.....kGy DOSE 29.7 - 37.9 <i>Validation Results.</i>
B2SAM0020044 B208090297	8045A-250S DISP SCOOP 25-45kGy	5	5 X B/No.196691 ACTUAL.....34.9.....kGy DOSE 30.7 - 36.5 <i>Validation Results.</i>
B2SAM0020045 B208090297	8045A-30S DISP PHARMA SCOOP 25-45kGy	3	3 X B/No.196692 CERTIFIED TO THE OUTSIDE OF THE BOX ONLY 35.8.....kGy
B2SAM002VAL	VALIDATION		SINGLE VALIDATION ON 0043 & 0044
	Total	224	Last Page 2 of 2

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Duce*

For and on behalf of ISOTRON LIMITED