

SAM002

Page 1

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT
Cust. Ref: 509-10386
Date Rec'd 18/02/09
Date: 26/02/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020008 B209020555	8055H-500S DISP VISCOTHIEF 25-45kGy	29	B/No.198686 ACTUAL....36.1.....kGy DOSE
B2SAM0020010 B209020555	8060H-510S DISP POWDER 25-45kGy	20	B/No.195147 ACTUAL....36.4.....kGy DOSE
B2SAM0020014 B209020555	8060H-1010S DISP POWDER 25-45kGy	20	B/No.198422 ACTUAL....36.5.....kGy DOSE
B2SAM0020016 B209020555	8050H-1000S DISP LIQUITHIEF 25-45kGy	30	B/No.198491 ACTUAL....35.3.....kGy DOSE
B2SAM0020020 B209020555	8060H-520S DISP POWDER 25-45kGy	10	B/No.198668 ACTUAL....35.6.....kGy DOSE
B2SAM0020021 B209020555	8095A-150S POWDER SPATULA 25-45kGy	80	B/No.196002

Continued... 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C. Drace*

For and on behalf of ISOTRON LIMITED

SAM002

Page 2

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT
Cust. Ref: 509-10386
Date Rec'd 18/02/09
Date: 26/02/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020025 B209020555	8095A-250 POWDER SPATULA 25-45kGy	28	ACTUAL....36.7 - 37.0....kGy DOSE B/No.198138 ACTUAL....36.5 - 36.8....kGy DOSE
B2SAM0020044 B209020555	8045A-250S DISP SCOOP 25-45kGy	12	B/No.196691 ACTUAL....36.3 - 37.2....kGy DOSE
B2SAM0020045 B209020555	8045A-30S DISP PHARMA SCOOP 25-45kGy	20	B/No.196692 ACTUAL....31.1 - 39.4....kGy DOSE - VALIDATION RESULTS SINGLE VALIDATION
B2SAM0020049 B209020555	8045A-125S DISP PHARM SCOOP 25-45kGy	6	B/No.196707 ACTUAL....37.3.....kGy DOSE
B2SAM002VAL	VALIDATION		
	Total	255	Last Page 2 of 2

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *C Drace*

For and on behalf of ISOTRON LIMITED