

SAM002

Page 1

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: 509-10887
Date Rec'd 16/09/09
Date: 23/09/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020008 B209090463	8055H-500S DISP VISCOTHIEF 25-45kGy	50	B/No.100382 ACTUAL.....35.1 - 35.8.....kGy DOSE
B2SAM0020016 B209090463	8050H-1000S DISP LIQUITHIEF 25-45kGy	60	B/No.100637 ACTUAL.....35.1 - 35.4.....kGy DOSE
B2SAM0020027 B209090463	8045A-1000S STERILE SCOOPS 25-45kGy	20	B/No.100386 ACTUAL.....35.7 - 35.9.....kGy DOSE
B2SAM0020038 B209090463	8045A-60S DISP PHARMA SCOOP 25-45kGy	20	B/No.198911 ACTUAL.....38.1.....kGy DOSE
B2SAM0020045 B209090463	8045A-30S DISP PHARMA SCOOP 25-45kGy	12	B/No.198807 ACTUAL.....36.2.....kGy DOSE
B2SAM0020063 B209090463	8063H-300S DIS POW LAN HDPE 25-45kGy	6	B/No.101972
			Continued.. 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED

SAM002

Page 2

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: 509-10887
Date Rec'd 16/09/09
Date: 23/09/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020052 B209090463	8045A-250S PHARMA SC00P 25-45kGy	8	MAX....38.0....MIN....29.3....kGy B/No.198900 ACTUAL....36.6 - 36.8....kGy DOSE
B2SAM002VAL	VALIDATION		3 X SINGLE VALIDATION ON B2SAM0020063 - 0064 - 0065
B2SAM0020064 B209090463	8450A-70S SCRAPER 70MM BLD 25-45kGy	4	B/No.09051 MAX....38.1....MIN....25.8....kGy
B2SAM0020065 B209090463	8450A-45S PAL KNIF 45MM BLD 25-45kGy	8	B/No.09051 MAX....35.9....MIN....27.0....kGy
Total		188	Last Page 2 of 2

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

For and on behalf of ISOTRON LIMITED