

SAM002

Page 1

Sampling Systems Ltd
 2 Forge Mills Park
 Station Road
 Coleshill
 Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

B46 1HT

Cust. Ref: S09-11106
 Date Rec'd 3/12/09
 Date: 11/12/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020008 B209120137	8055H-500S DISP VISCOTHIEF 25-45kGy	30	30 X B/No.103161 ACTUAL DOSE....34.0.....kGy
B2SAM0020014 B209120137	8060H-1010S DISP POWDER 25-45kGy	30	30 X B/No.101892 ACTUAL DOSE....33.9.....kGy
B2SAM0020016 B209120137	8050H-1000S DISP LIQUITHIEF 25-45kGy	30	30 X B/No.101888 ACTUAL DOSE....33.8.....kGy
B2SAM0020019 B209120137	8060H-1020S DISP POWDER 25-45kGy	37	37 X B/No. 101897 ACTUAL DOSE....33.5.....kGy
B2SAM0020020 B209120137	8060H-520S DISP POWDER 25-45kGy	63	63 X B/No.101889 ACTUAL DOSE....34.3.....kGy
B2SAM0020027 B209120137	8045A-1000S STERILE SC00PS 25-45kGy	20	20 X B/No.102803 ACTUAL DOSE....34.3 - 35.1....kGy
B2SAM0020047 B209120137	8073H-1000S M/LEVEL SAMPLER 25-45kGy	25	25 X B/No.101894

Continued.. 1

This is to certify that the above items have been irradiated as specified above.

Authorising Signature: *J. Goffey*

For and on behalf of ISOTRON LIMITED



SAM002

Page 2

Sampling Systems Ltd
 2 Forge Mills Park
 Station Road
 Coleshill
 Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

B46 1HT
 Cust. Ref: S09-11106
 Date Rec'd 3/12/09
 Date: 11/12/09

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020048 B209120137	8073H-500S M/LEVEL SAMPLER 25-45kGy	42	ACTUAL DOSE....32.6 - 32.8....kGy 42 X B/No.199943 ACTUAL DOSE....32.4 - 33.0....kGy
B2SAM0020067 B209120137	8063H-300S DIS POWDER LAN 25-45kGy	12	12 X B/No.102534 ACTUAL DOSE....34.6.....kGy
B2SAM002VAL	VALIDATION		SINGLE VALIDATION ON 8063H-300S DISP POWDER LANCES B2SAM0020067
Total		289	Last Page 2 of 2

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED