

SAM002

Page 1

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: S10-11173
Date Rec'd 23/12/09
Date: 11/01/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020008 B210010125	8055H-500S DISP VISCOTHIEF 25-45kGy	10	10 X B/No.102522 ACTUAL DOSE....34.7.....kGy
B2SAM0020010 B210010125	8060H-510S DISP POWDER 25-45kGy	10	10 X B/No.102516 ACTUAL DOSE....34.7.....kGy
B2SAM0020015 B210010125	8050H-500S DISP LIQUITHIEF 25-45kGy	10	10 X B/No.102521 ACTUAL DOSE....34.7.....kGy
B2SAM0020020 B210010125	8060H-520S DISP POWDER 25-45kGy	10	10 X B/No.102518 ACTUAL DOSE....34.7.....kGy
B2SAM0020024 B210010125	8063H-1000S DISP POWDER 25-45kGy	10	10 X B/No.102524 ACTUAL DOSE....34.5.....kGy
B2SAM0020027 B210010125	8045A-1000S STERILE SCOOPS 25-45kGy	126	54 X B/No.103182 72 X B/No.103301 ACTUAL DOSE....34.8 - 35.3....kGy
B2SAM0020033	8045A-500S DIS PHARM SCOOPS 25-45kGy	54	54 X B/No.103383

Continued.. 1

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED



SAM002

Page 2

Sampling Systems Ltd
 2 Forge Mills Park
 Station Road
 Coleshill
 Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
 Bradford · West Yorkshire
 BD4 6SE · United Kingdom
 Telephone: +44 (0)1274 686011
 Facsimile: +44 (0)1274 686061
 Email: assistance@isotron.com
 Website: www.isotron.com

B46 1HT

Cust. Ref: S10-11173
 Date Rec'd 23/12/09
 Date: 11/01/10

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B210010125			ACTUAL DOSE....36.8 - 37.5....kGy
B2SAM0020038 B210010125	8045A-60S DISP PHARMA SCOOP 25-45kGy	15	15 X B/No.103375 ACTUAL DOSE....35.7.....kGy
B2SAM0020059 B210010125	8045R-1000S STERILE SCOOPS 25-45kGy	54	54 X B/No.103591 ACTUAL DOSE....35.6 - 36.4....kGy
B2SAM0020068 B210010125	8370A-100S FUNNEL 100MM 25-45kGy	10	10 X B/No.103346 ACTUAL DOSE....35.8.....kGy
B2SAM002VAL	VALIDATION		SINGLE VALIDATION ON B2SAM0020068 VALIDATION REF: 4.0705
B2SAM0020023 B210010125	8063H-500S DISP POWDER 25-45kGy	10	B/No.102526 ACTUAL DOSE....34.2.....kGy
	Despatch Text: 21PLTS		
	Total	319	Last Page 2 of 2

This is to certify that the above items have been irradiated as specified above

Authorising Signature: 

