

SAM002

Page 1

Sampling Systems Ltd  
2 Forge Mills Park  
Station Road  
Coleshill  
Warwickshire



**Isotron Limited**  
Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

B46 1HT

S11-12244 & S11-12253

Cust. Ref:

5/01/11

Date Rec'd

Date: 10/01/11

| ITEM CODE<br>ISOTRON BATCH | ITEM SPECIFICATION                   | QTY | ADDITIONAL DETAILS                            |
|----------------------------|--------------------------------------|-----|---|
| B2SAM0020010<br>B211010157 | 8060H-510S DISP POWDER 25-45kGy      | 13  | B/No.110091<br><br>ACTUAL DOSE..36.6...kGy    |
| B2SAM0020019<br>B211010157 | 8060H-1020S DISP POWDER 25-45kGy     | 47  | B/No.110309<br><br>ACTUAL DOSE.34.5-34.9..kGy |
| B2SAM0020020<br>B211010157 | 8060H-520S DISP POWDER 25-45kGy      | 105 | B/No.110086<br><br>ACTUAL DOSE.35.0-35.6..kGy |
| B2SAM0020025<br>B211010157 | 8095A-250 POWDER SPATULA 25-45kGy    | 21  | B/No.107211<br><br>ACTUAL DOSE.35.7-36.1..kGy |
| B2SAM0020038<br>B211010157 | 8045A-60S DISP PHARMA SCOOP 25-45kGy | 31  | B/No.109030<br><br>ACTUAL DOSE.37.1..kGy      |
| B2SAM0020042<br>B211010157 | 8095A-150S POWDER SPATULA 25-45kGy   | 30  | B/No.109061<br><br>ACTUAL DOSE.36.3-37.0..kGy |
| B2SAM0020043<br>B211010157 | 8045A-2500S DISP SCOOP 25-45kGy      | 120 | B/No.109022                                   |
|                            |                                      |     | Continued.. 1                                 |

**EXPIRY DATE IS 2 YEARS  
FROM DATE OF IRRADIATION  
PROVIDED THE BAG REMAINS INTACT**

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

**CERTIFIED FINAL**

For and on behalf of ISOTRON LIMITED

SAM002

Page 2

Sampling Systems Ltd  
2 Forge Mills Park  
Station Road  
Coleshill  
Warwickshire



**Isotron Limited**  
Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

B46 1HT  
S11-12244 & S11-12253

Cust. Ref: S11-12244 & S11-12253  
Date Rec'd 5/01/11  
Date: 10/01/11

| ITEM CODE<br>ISOTRON BATCH | ITEM SPECIFICATION                  | QTY | ADDITIONAL DETAILS  |
|----------------------------|-------------------------------------|-----|---|
|                            |                                     |     | ACTUAL DOSE .37.2-37.3..kGy   |
| B2SAM0020048<br>B211010157 | 8073H-500S M/LEVEL SAMPLER 25-45kGy | 29  | B/No.109028<br><br>ACTUAL DOSE .35.9..kGy   |
| B2SAM0020059<br>B211010157 | 8045R-1000S STERILE SCOOPS 25-45kGy | 56  | B/No.108775<br><br>ACTUAL DOSE .35.8-37.2..kGy  |
| B2SAM0020069<br>B211010157 | 8049H-2000S 2LTR JUG 25-45kGy       | 10  | B/No.108490<br><br>ACTUAL DOSE .36.8-37.1..kGy  |
| B2SAM0020080<br>B211010157 | 5393-1131 LIQ DISPO 190ML 25-45kGy  | 10  | B/No.110150<br><br>ACTUAL DOSE .36.0..kGy   |
| B2SAM0020086<br>B211010157 | 8050H-500S DIS LIQUITHIEF 25-45kGy  | 45  | B/No.109039<br><br>CERTIFIED TO THE EXTERNAL DOSIMETER<br>READING ONLY<br><br>.35.0-35.8..kGy |
| B2SAM0020088               | 8640H-10S VOL SPOON 10ML 25-45kGy   | 10  | B/No.109905<br><br>Continued.. 2  |

**EXPIRY DATE IS 2 YEARS  
FROM DATE OF IRRADIATION  
PROVIDED THE BAG REMAINS INTACT**

This is to certify that the above items have been irradiated as specified above

Authorising Signature: 



For and on behalf of ISOTRON LIMITED

SAM002

Page 3

Sampling Systems Ltd  
2 Forge Mills Park  
Station Road  
Coleshill  
Warwickshire



**Isotron Limited**  
Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@isotron.com  
Website: www.isotron.com

B46 1HT

S11-12244 & S11-12253

Cust. Ref:

Date Rec'd 5/01/11

Date: 10/01/11

| ITEM CODE<br>ISOTRON BATCH | ITEM SPECIFICATION                | QTY | ADDITIONAL DETAILS                           |
|----------------------------|-----------------------------------|-----|--|
| B211010157                 |                                   |     | ACTUAL DOSE .36.1..kGy                       |
| B2SAM0020089<br>B211010157 | 8640H-20S VOL SPOON 20ML 25-45kGy | 10  | B/No .109906<br>ACTUAL DOSE .36.5..kGy       |
| B2SAM002VAL                | VALIDATION                        |     | SINGLE VALIDATION ON 8640H-10S<br>Ref:4.0890 |
|                            | Despatch Text:<br><br>22PLTS      |     |  |
|                            | Total                             | 537 | Last Page 3 of 3                             |

**EXPIRY DATE IS 2 YEARS  
FROM DATE OF IRRADIATION  
PROVIDED THE BAG REMAINS INTACT**

This is to certify that the above items have been irradiated as specified above

Authorising Signature:

**CERTIFIED FINAL**

For and on behalf of ISOTRON LIMITED