

CERTIFICATE OF IRRADIATION Number: **B** 0476738 00

SAM002

Page 1

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited
Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT
Cust. Ref: S11-12321
Date Rec'd 28/01/11
Date: 7/02/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020019 B211020037	8060H-1020S DISP POWDER 25-45kGy	55	B/No.110308 ACTUAL DOSE....32.4 - 34.1....kGy
B2SAM0020027 B211020040	8045A-1000S STERILE SCOOPS 25-45kGy	60	B/No.109013 ACTUAL DOSE....34.7 - 38.3....kGy
B2SAM0020031 B211020038	8097A-150S POWDER SAFE 25-45kGy	5	B/No.195149 ACTUAL DOSE....34.9....kGy
B2SAM0020049 B211020040	8045A-125S DISP PHARM SCOOP 25-45kGy	12	B/No.109015 ACTUAL DOSE....34.7 - 38.3....kGy
B2SAM0020076 B211020037	8055H-1000S DIS VIS 25-45kGy	120	B/No.109045 ACTUAL DOSE....34.1....kGy
B2SAM0020077 B211020037	8060H-1005S DISP POWDER 25-45kGy	12	B/No.110092 ACTUAL DOSE....32.5....kGy
B2SAM0020086 B211020041	8050H-500S DIS LIQUITHIEF 25-45kGy	30	B/No.110244
			Continued.. 1

**EXPIRY DATE IS 2 YEARS
FROM DATE OF IRRADIATION
PROVIDED THE BAG REMAINS INTACT**

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Coffey*

For and on behalf of ISOTRON LIMITED

CERTIFICATE OF IRRADIATION

Number: **B** 0476738 00

SAM002

Page 2

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: S11-12321
Date Rec'd 28/01/11
Date: 7/02/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020087 B211020041	8055H-500S DIS VISCOTHIEF 25-45kGy	40	ACTUAL DOSE....34.1 - 37.8....kGy B/No.109047
B2SAM0020090 B211020041	8088A-10S SAMPLE SPOON 10ML 25-45kGy	30	ACTUAL DOSE....34.1 - 37.8....kGy B/No.109910
B2SAM0020091 B211020041	8088A-03S SAMPLE SPOON 3ML 25-45kGy	30	ACTUAL DOSE....34.4 - 35.1....kGy B/No.109908
B2SAM0020092 B211020041	8640H-01S VOLM SPOON 1ML 25-45kGy	11	ACTUAL DOSE....34.4 - 35.1....kGy B/No.109903
B2SAM0020093 B211020041	8640H-05S VOLM SPOON 5ML 25-45kGy	10	ACTUAL DOSE....34.4 - 35.1....kGy B/No.109904
B2SAM0020094 B211020039	8015N-170S MICRO SPATULA 25-45kGy	40	ACTUAL DOSE....34.5....kGy B/No.110151
			Continued.. 2

EXPIRY DATE IS 2 YEARS
FROM DATE OF IRRADIATION
PROVIDED THE BAG REMAINS INTACT

This is to certify that the above items have been irradiated as specified above

Authorising Signature

CERTIFIED FINAL

For and on behalf of ISOTRON LIMITED

CERTIFICATE OF IRRADIATION

Number: **B** 0476738 00

SAM002

Page 3

Sampling Systems Ltd
2 Forge Mills Park
Station Road
Coleshill
Warwickshire



Isotron Limited

Roydsdale Way · Euroway Trading Estate
Bradford · West Yorkshire
BD4 6SE · United Kingdom
Telephone: +44 (0)1274 686011
Facsimile: +44 (0)1274 686061
Email: assistance@isotron.com
Website: www.isotron.com

B46 1HT

Cust. Ref: 511-12321
Date Rec'd 28/01/11
Date: 7/02/11

ITEM CODE ISOTRON BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
	Despatch Text: 15PLTS		
	Total	455	Last Page 3 of 3

EXPIRY DATE IS 2 YEARS
 FROM DATE OF IRRADIATION
 PROVIDED THE BAG REMAINS INTACT

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *J. Giffen*

CERTIFIED FINAL

For and on behalf of ISOTRON LIMITED