

SAM002

Page 1

Sampling Systems Ltd  
2 Forge Mills Park  
Station Road  
Coleshill  
Warwickshire



**Synergy Health plc**  
Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: assistance@synergyhealthplc.com  
Website: www.synergyhealthplc.com

B46 1HT  
Cust. Ref: S12-13162  
Date Rec'd 9/01/12  
Date: 11/01/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
B2SAM0020023 B212010166	8063H-500S DISP POWDER 25-45kGy	45	B/No.5551  ROUTINE DOSE 34.3, 34.6 kGy
B2SAM0020024 B212010166	8063H-1000S DISP POWDER 25-45kGy	30	B/No.5549  ROUTINE DOSE 34.2, 34.7 kGy
B2SAM0020038 B212010166	8045A-60S DISP PHARMA SCOOP 25-45kGy	41	B/No.6510  ROUTINE DOSE 35.8, 35.6 kGy
B2SAM0020042 B212010166	8095A-150S POWDER SPATULA 25-45kGy	41	B/No.6917  ROUTINE DOSE 35.0, 35.0 kGy
B2SAM0020047 B212010166	8073H-1000S M/LEVEL SAMPLER 25-45kGy	20	B/Bo.5712  ROUTINE DOSE 33.1 kGy
B2SAM0020048 B212010166	8073H-500S M/LEVEL SAMPLER 25-45kGy	20	B/No.5555  ROUTINE DOSE 32.8 kGy
B2SAM0020069 B212010166	8049H-2000S 2LTR JUG 25-45kGy	30	B/No.5537
			Continued.. 1

**EXPIRY DATE IS 2 YEARS  
FROM DATE OF IRRADIATION  
PROVIDED THE BAG REMAINS INTACT**

This is to certify that the above items have been irradiated as specified above

Authorising Signature: *[Signature]*

For and on behalf of SYNERGY HEALTH PLC



# CERTIFICATE OF IRRADIATION

Number: **B** 0516296 00

SAM002

Page 2

Sampling Systems Ltd  
2 Forge Mills Park  
Station Road  
Coleshill  
Warwickshire



## Synergy Health plc

Roydsdale Way · Euroway Trading Estate  
Bradford · West Yorkshire  
BD4 6SE · United Kingdom  
Telephone: +44 (0)1274 686011  
Facsimile: +44 (0)1274 686061  
Email: [assistance@synergyhealthplc.com](mailto:assistance@synergyhealthplc.com)  
Website: [www.synergyhealthplc.com](http://www.synergyhealthplc.com)

B46 1HT  
Cust. Ref: S12-13162  
Date Rec'd 9/01/12  
Date: 11/01/12

ITEM CODE SYNERGY HEALTH BATCH	ITEM SPECIFICATION	QTY	ADDITIONAL DETAILS
			ROUTINE DOSE 35.6, 35.6, 36.0 kGy
B2SAM0020087 B212010166	8055H-500S DIS VISCOTHIEF 25-45kGy	40	B/No.002662  ROUTINE DOSE 34.9, 34.9 kGy
B2SAM0020098 B212010166	8051H-450S JUMBO PIPETTE 25-45kGy	30	B/No.5543  ROUTINE DOSE 33.3, 33.2 kGy
B2SAM0020113 B212010166	8050H-1000S DIS LIQ 25-45kGy	75	40 X B/No.001015 35 X B/No.5540
			ROUTINE DOSE 34.3, 34.8, 34.8 kGy
	Despatch Text:  15PLTS		
	Total	372	Last Page 2 of 2

EXPIRY DATE IS 2 YEARS  
 FROM DATE OF IRRADIATION  
 PROVIDED THE BAG REMAINS INTACT

This is to certify that the above items have been irradiated as specified above

Authorising Signature:



For and on behalf of SYNERGY HEALTH PLC